


















**New Parts to fit Midmark® • Ritter® Sterilizers**





**M7 & 7 SpeedClave®**

**MIDMARK® • RITTER®**

 	 	 
<p><b>RPI Part #RCV112</b>  <b>OEM Part #014-0236-00</b>  <b>SOLENOID VALVE (FILL/VENT)</b>  <b>Fits:</b> Mounts on Chassis  <b>Models:</b> M7-020, -022                  All Serial Numbers with V Prefix</p>	<p><b>RPI Part #RCK113</b>  <b>OEM Part #014-0236-11</b>  <b>REPAIR KIT (FILL/VENT)</b>  <b>Fits:</b> Solenoid Valve (Fill/Vent)  <b>Models:</b> M7-020, -022                  All Serial Numbers with V Prefix</p>	<p><b>RPI Part #RCC114</b>  <b>OEM Part #014-0236-10</b>  <b>COIL (FILL/VENT)</b>  <b>Fits:</b> Solenoid Valve (Fill/Vent)  <b>Models:</b> M7-020, -022                  All Serial Numbers with V Prefix</p>
<p><b>Need Repair Kit?</b> →</p>		<p>→ <b>Need Coil?</b></p>

 		
<p><b>RPI Part #RCS115</b>  <b>OEM Part #015-1791-00</b>  <b>FILL/VENT SWITCH</b>  <b>Fits:</b> Front Panel  <b>Models:</b> M7-020 thru -022                  All Serial Numbers with V Prefix</p>	<p><b>RPI Part #RPT501</b>  <b>OEM Part # (No OEM # Available)</b>  <b>DOUBLE ENDED SPANNER WRENCH</b>  <b>Fits:</b> Solenoid Valve (Fill/Vent)  <b>Models:</b> M7-020 thru -022                  All Serial Numbers with V Prefix</p>	<p><b>RPI Part #ILS014</b>  <b>OEM Part # (No OEM # Available)</b>  <b>SPANNER NUT</b>  <b>Fits:</b> Solenoid Valve (Fill/Vent)  <b>Models:</b> M7-020 thru -022                  All Serial Numbers with V Prefix</p>

	 	 
<p><b>RPI Part #MIF062</b>  <b>New OEM Part #002-0360-00</b>  <b>Old OEM Part #016-0790-00</b>  <b>FILL/VENT MESH CHAMBER FILTER (Stainless Steel)</b>  <b>Fits:</b> Chamber  <b>Models:</b> M7-020 thru -022                  All Serial Numbers with V Prefix</p>	<p><b>RPI Part #RPH674</b>  <b>OEM Part #015-1259-04</b>  <b>FUSE HOLDER</b>  <b>Fits:</b> Rear Panel  <b>Models:</b> M7-020, -022                  All Serial Numbers with V Prefix</p>	<p><b>RPI Part #RPF679</b>  <b>OEM Part #015-0346-35</b>  <b>FUSE (12A, 250VAC) - FAST ACTING (5/pkg)</b>  <b>Fits:</b> Fuse Holder (RPI Part #RPH674)  <b>Models:</b> M7-020, -022                  All Serial Numbers with V Prefix</p>


			
<p><b>RPI Part #PCC003</b>  <b>OEM Part #H98159</b>  <b>INDUSTRIAL GRADE POWER CORD</b>  <b>Fits:</b> Lower Chassis  <b>Models:</b> M7 &amp; 7 SpeedClave® All Models</p>	<p><b>RPI Part #RPB167</b>  <b>OEM Part #H98160</b>  <b>STRAIN RELIEF BUSHING (6/pkg)</b>  <b>Fits:</b> Power Cord on Lower Chassis  <b>Models:</b> M7 &amp; 7 SpeedClave® All Models</p>	<p><b>RPI Part #RPT172</b>  <b>OEM Part # (No OEM # Available)</b>  <b>STRAIN RELIEF BUSHING PLIERS</b>  <b>Fits:</b> Strain Relief Bushing  <b>Models:</b> M7 &amp; 7 SpeedClave® All Models</p>	

# Table Top Sterilizers

## M7 & 7 SpeedClave®

RPI Part #RCB088  
OEM Part #H96772

**T-BOLT**  
Fits: Door Assembly  
Models: M7 & 7 SpeedClave®  
All Models



RPI Part #RCG104  
OEM Part #053-1579-00

**FIBER GASKET** (5/pkg)  
Fits: Door Assembly  
Models: M7 & 7 SpeedClave®  
All Models



RPI Part #RCW009  
OEM Part #H96771 & 3059

**DOOR WASHER**  
Fits: T-Bolt on Door Assembly  
Models: M7 & 7 SpeedClave®  
All Models




RPI Part #RCN094  
OEM Part #H96773

**CAP NUT**  
Fits: T-Bolt in Door Assembly  
Models: M7 & 7 SpeedClave®  
All Models



RPI Part #RPA032  
OEM Part #042-0024-02

**THREADLOCKER 242** (5/pkg)  
Fits: Cap Nut and T-Bolt on the Door Assembly  
Models: M7




RPI Part #RCS108  
OEM Part #H106330

**DOOR SPRING** (5/pkg)  
Fits: Door Assembly  
Models: M7 & 7 SpeedClave®  
All Models




RPI Part #RPK431  
OEM Part #002-0653-00

**DRAIN TUBE ASSEMBLY KIT**  
Fits: Front Panel  
Models: M7 (-011 thru -016) & (-020 thru -022)



RPI Part #RPF427  
OEM Part #053-0990-01

**MALE QUICK CONNECT FITTING (GREY)**  
Fits: Drain Tube Assembly Kit (RPI Part #RPK431)  
Models: M7 (-011 thru -016) & (-020 thru -022)



RPI Part #RPO304  
OEM Part # (No OEM # Available)  
Industry Part #AS568-011


**O-RING** (12/pkg)  
Fits: Male Quick Connect Fitting (Grey) (RPI Part #RPF427)  
Models: M7 (-011 thru -016) & (-020 thru -022)



SHOWN  
ACTUAL  
SIZE

RPI Part #RPF426  
OEM Part #053-0990-00

**FEMALE QUICK CONNECT FITTING (GREY)**  
Fits: Front Panel  
Models: M7 (-011 thru -016) & (-020 thru -022)




RPI Part #RPH673  
OEM Part #040-0010-125

**SELF DRILLING SCREW** (6/pkg)  
Fits: Cover  
Models: M7 & 7 SpeedClave®  
All Models



RPI Part #RPH285  
OEM Part #053-0391-03 & 053-0391-02

**DRAIN HOSE**  
Fits: Fits: Male Quick Connect Fitting (Grey) (RPI Part #RPF427)  
Models: M7 (-011 thru -016) & (-020 thru -022)



Sold by the foot; Reel not included

RPI Part #RPC286  
OEM Part # (No OEM # Available)

**KWIK™ CLAMP** (6/pkg)  
Fits: Drain Hose/Inner Tubing  
Models: M7 (-011 thru -016) & (-020 thru -022)



### SERVICE TIP


The efficiency of a sterilizer itself should be monitored periodically to determine if proper sterilization is taking place. And one of the ways to do this is to check that the sterilizer is performing at the proper heat temperature.

A good tool to check the temperature inside a sterilizer is the RPI Max Register Thermometer. This thermometer registers the highest temperature it has been exposed to and holds that temperature until reset. To reset, simply shake it until the mercury inside reaches ambient temperature.

For more ways to monitor a sterilizer for optimal performance, please see the "Tech Talk" article in our September, 2008 newsletter.

RPI Part #RPT113  
OEM Part # (No OEM # Available)

**MAX REGISTER THERMOMETER**  
• Use to test temperature inside sterilizer



Look for more parts to fit M7 & 7 SpeedClave Sterilizers coming soon!